THERAFLEX UV PLATELETS: NOTHING BUT UVC LIGHT AND STRONG AGITATION

H. Mohr[1], U. Gravemann[1], F. Tolksdorf[2], W.H. Walker[2], T.H. Müller[1]

Purpose

Blood donations may not only be contaminated with viruses, e.g. HBV, HCV or HIV. In addition, they may contain bacteria. This is especially crucial for platelet concentrates (PCs), because they have to be stored at room temperature, at which bacteria can multiply to high levels [1-2].

Short-wave ultraviolet light (UVC, wavelength range: 200-280 nm) is germicidal, but low UV-permeability hampers its use for sterilizing PCs. A simple method was developed which overcomes this limitation.

Materials and Methods

Plasma-reduced PCs in storage medium SSP+ (volume approx. 350 mL, platelet concentration approx. 10^9/mL, plasma content 30-35%) were prepared from pools of 5 buffy coats [3]. PC volume was approx. 350 mL. The PCs were spiked with approx. 10^2-10^6 CFU/mL of different bacteria species or up to 10^7 $TCID_{50}$/mL of lipid-enveloped or nonenveloped viruses. Other PCs were spiked with $5x10^6$/mL peripheral blood mononuclear cells (PBMC). The PCs were filled into UV-transparent plastic bags and irradiated on a device (Fig.1), equipped with mercury vapour tubes emitting monochromatic UVC-light (wavelength: 254 nm). The device was equipped with an orbital agitator. Irradiation was from both sides of the bags. UVC doses applied were up to 0.6 J/cm² (approx. 90 sec). During treatment the PCs were strongly agitated. Bacteria or virus titers, PBMC viability and platelet parameters were determined before and after irradiation. Each experiment was repeated 3-6 times. Results are depicted as mean ± SD.

Results

Pathogen inactivation was enormously enhanced when the PCs were loosely placed on a quartz plate located between the two layers of UVC tubes of the irradiation device and, in addition, strongly agitated during irradiation (Fig. 2).

UVC-light at 0.3-0.4 J/cm² (irradiation time: approx. 1 min) reduced the titers of all bacteria tested by approx. 5-6 log_{10} steps. PCs spiked with approx. 100 CFU/ml of bacteria were reproducibly sterilized (Tab.1). In one experiment with *B. cereus* the PC was sterile after 3 but unsterile after 6 days storage. This was probably due to spores of *B. cereus* that are more resistant to UVC than vegetative bacteria.

UVC sensitivity of the viruses tested was not so uniform (Table 1): The small single stranded RNA viruses VSV, Sindbis and WNV were completely inactivated at approx. 0.3-0.4 J/cm². Remarkably HIV-1 (also a small single-stranded RNA virus) was only moderately inactivated at UVC doses up to 0.6 J/cm².

The small nonenveloped DNA viruses PPV and EMCV proved to be very sensitive. Complete inactivation was achieved at 0.4-0.5 J/cm².

With the exception of HIV-1, SHV-1 was more resistant than the other viruses tested. This confirms that in general large double stranded DNA viruses are not as susceptible to UVC as smaller single stranded DNA or RNA viruses.

PBMC proved to be extremely sensitive to UVC irradiation: Complete inactivation was found at less than 0.1 J/cm² (Fig. 3)

PC properties remained almost unchanged at doses up to 0.6 J/cm². The storage stability of the treated PCs for up to 6 days after treatment (8 days after blood donation) was maintained (Table 2)

Conclusions

Irradiation with UVC under strong agitation may be used to sterilize platelet concentrates at a light dose that is not harmful to the products. The UVC dose required is 0.4 J/cm². Irradiation time is not more than approx. 1 min.

Parameter	Day 1 after irradiation				Day 6 after irradiation			
	Control	UVC dose (J/cm²)			Control	UVC dose (J/cm²)		
		0.4	0.5	0.6		0.4	0.5	0.6
Pts [x10^9/mL]	10.8 ± 0.6	10.2 ± 0.6	9.8 ± 0.6	9.1 ± 0.9	10.1 ± 0.8	9.8 ± 0.6	9.3 ± 0.8	9.3 ± 0.9
pH	7.10 ± 0.04	7.04 ± 0.05	7.09 ± 0.05	7.05 ± 0.04	7.27 ± 0.15	7.09 ± 0.06	7.11 ± 0.10	6.98 ± 0.07
Lactate [mmol/L]	7.7 ± 1.0	8.0 ± 0.5	7.7 ± 0.5	8.0 ± 0.7	12.7 ± 1.0	14.9 ± 1.0	14.6 ± 1.4	16.7 ± 1.4
Glucose [mg/dL]	122 ± 9	117 ± 7	117 ± 6	115 ± 7	62 ± 11	43 ± 8	44 ± 11	29 ± 10
Swirling	ok	ok	ok	ok	ok	ok	ok	ok
HSR [%]	69 ± 5	66 ± 2	61 ± 6	62 ± 4	68 ± 2	65 ± 2	62 ± 3	56 ± 5
Collagen-induced aggregation [%]	95 ± 4	90 ± 5	88 ± 3	87 ± 2	62 ± 9	69 ± 8	67 ± 2	69 ± 5
CD62 [%]	36 ± 1	46 ± 3	47 ± 2	49 ± 1	29 ± 1	45 ± 8	50 ± 10	57 ± 8
Annexin V [%]	5 ± 1	6 ± 3	7 ± 4	7 ± 4	9 ± 5	8 ± 2	10 ± 2	12 ± 3

Tab. 3: Treatment of PCs with different UVC doses. Influence on platelet parameters and on storage stability. n=6, mean + SD



***Fig. 1*: Irradiation device for UVC treatment of PCs**





***Fig. 2*: Inactivation of *St. epidermidis* in PC aliquots (110 or 280 mL) by irradiation with UV light: fixed vs. loosely placed irradiation bags (a); dependence of bacteria inactivation in loosely placed irradiation bags on the agitation speed (b). n=3, mean ± SD**

Bacteria species	Characteristics	Gram stain	Number pf experiments	Spike (CFU/mL)	BacT/Alert result*	Remark
B. cereus	fac. anaerobic	pos	12	100-140	11 sterile 1 unsterile**	Spore former
E. coli	aerobic	neg	12	36-65	12 sterile	
K. pneumoniae	fac. anaerobic	neg	12	85-140	12 sterile	
P. acnes	anaerobic	neg	12	61-100	12 sterile	
S. aureus	fac. anaerobic	pos	22	60-110	22 sterile	
S. epidermidis	fac. anaerobic	pos	22	74-210	22 sterile	
Str. pyogenes	fac. anaerobic	pos	12	118-194	12 sterile	

*: Samples (2x10 mL each) were drawn after 3 and 6 days at 22°C
**: sterile after 3 days storage

***Tab 1:* Sterilization of PCs spiked with different bacteria species by irradiation with UVC (0.4 J/cm²)**

Virus	Genome	Lipid Envelope	Model virus for	Log_{10}reduction factor
Vesicular stomatitis (VSV)	ss* RNA	X	-	≥ 6.41
Sindbis (Sindbis)	ss RNA	X	-	5.55
West Nile (WNV)	ss RNA	X	HCV	5.24
Human Immunodeficiency (HIV-1)	ss RNA	X	-	1.36
Suid Herpes (SHV-1)	ds** DNA	X	HBV/CMV	3.57
Porcine Parvo (PPV)	ss DNA	-	Parvo B 19	≥ 6.42
Encephalomyocarditis (EMCV)	ss DNA	-	HAV	5.73

Tab 2: Inactivation factors of viruses by irradiation with UVC (0.4 J/cm²)



Fig. 3: Inactivation of T-lymphocytes in platelet concentrates by irradiation with UVC. Viability was assayed by mixed lymphocyte culture.

References

1. Wagner SJ. Transfusion-transmitted bacterial infection: risks, sources and interventions. Vox Sang 2004;86(3):157-63.
2. Mohr H, Bayer A, Gravemann U, Muller TH. Elimination and multiplication of bacteria during preparation and storage of buffy coat-derived platelet concentrates. Transfusion 2006;46(6):949-55.
3. Eriksson L, Shanwell A, Gulliksson H, et al. Platelet concentrates in an additive solution prepared from pooled buffy coats. In vivo studies. Vox Sang 1993;64(3):133-8

[1] Blood Center of the German Red Cross Chapters of NSTOB, Springe, Germany
[2] MacoPharma Int., Langen, Germany

Supported by Forschungsgemeinschaft der Blutspendedienste des Deutschen Roten Kreuzes e.V.